

# (12) UK Patent Application (19) GB (11) 2 327 351 (13) A

(43) Date of A Publication 27.01.1999

(21) Application No 9718199.7

(22) Date of Filing 18.07.1997

(71) Applicant(s)  
Gyrus Medical Limited  
(Incorporated in the United Kingdom)  
Fortran Road, St Mellons, CARDIFF, CF3 0LT,  
United Kingdom

(72) Inventor(s)  
Nigel Mark Goble  
Colin Charles Owen Goble

(74) Agent and/or Address for Service  
Withers & Rogers  
4 Dyer's Buildings, Holborn, LONDON, EC1N 2JT,  
United Kingdom

(51) INT CL<sup>6</sup>  
A61B 17/39

(52) UK CL (Edition Q )  
A5R RHCC RHCE

(56) Documents Cited  
GB 2308981 A EP 0754437 A1

(58) Field of Search  
UK CL (Edition O ) A5R RHCC RHCE  
INT CL<sup>6</sup> A61B 17/36 17/38 17/39  
Online: EDOC, JAPIO, WPI

(54) Abstract Title  
Electrosurgical instrument

(57) An electrosurgical instrument for the treatment of tissue in the presence of an electrically conductive fluid medium comprises an instrument shaft 10, and an active side-effect tissue treatment electrode 12 at the distal end of a tungsten rod 14 within the shaft 10 for electrosurgically excising tissue pieces at an operation site. A ceramic tip 20 has an aperture 20a through which matter can be aspirated and removed from the region surrounding the electrode 12 by way of a channel formed within the shaft 10 and leading from the aperture 20a to a suction port 10a. The tissue treatment electrode 12 is coupled to a motor 44 so as to move cyclically relative to the shaft 10, and causes electrosurgical morcellation of the tissue pieces. Vaporisation of the fluid medium and tissue is controlled by pulsing the suction pump or adjusting the turbulence produced by reciprocating the electrode 12. Vaporisation is detected by a electrode peak voltage sensor. An RF power source is coupled to connectors 46,48. The rod 14 may be offset or cranked, and the electrode 12 may be end-effect.




Fig.2.

2327351

1

AN ELECTROSURGICAL INSTRUMENT

This invention relates to an electrosurgical instrument for the treatment of tissue in the  
5 presence of an electrically-conductive fluid medium, to electrosurgical apparatus  
including such an instrument, and to an electrode unit for use in such an instrument.  
Endoscopic electrosurgery is useful for treating tissue in cavities of the body, and is  
normally performed in the presence of a distension medium. When the distension  
medium is a liquid, this is commonly referred to as underwater electrosurgery, this term  
10 denoting electrosurgery in which living tissue is treated using an electrosurgical  
instrument with a treatment electrode or electrodes immersed in liquid at the operation  
site. A gaseous medium is commonly employed when endoscopic surgery is performed  
in a distensible body cavity of larger potential volume in which a liquid medium would  
be unsuitable, as is often the case in laparoscopic or gastroenterological surgery.

15 Underwater surgery is commonly performed using endoscopic techniques, in which the  
endoscope itself may provide a conduit (commonly referred to as a working channel)  
for the passage of an electrode. Alternatively, the endoscope may be specifically  
adapted (as in a resectoscope) to include means for mounting an electrode, or the  
20 electrode may be introduced into a body cavity via a separate access means at an angle  
with respect to the endoscope - a technique commonly referred to as triangulation.  
These variations in technique can be subdivided by surgical speciality, where one or  
other of the techniques has particular advantages given the access route to the specific  
body cavity. Endoscopes with integral working channels, or those characterised as  
25 resectoscopes, are generally employed when the body cavity may be accessed through  
a natural body opening - such as the cervical canal to access the endometrial cavity of  
the uterus, or the urethra to access the prostate gland and the bladder. Endoscopes  
specifically designed for use in the endometrial cavity are referred to as hysteroscopes,  
and those designed for use in the urinary tract include cystoscopes, urethrosopes and  
30 resectoscopes. The procedures of transurethral resection or vaporisation of the prostate  
gland are known as TURP and EVAP respectively. When there is no natural body

opening through which an endoscope may be passed, the technique of triangulation is commonly employed. Triangulation is commonly used during underwater endoscopic surgery on joint cavities such as the knee and the shoulder. The endoscope used in these procedures is commonly referred to as an arthroscope.

5

Electrosurgery is usually carried out using either a monopolar instrument or a bipolar instrument. With monopolar electrosurgery, an active electrode is used in the operating region, and a conductive return plate is secured to the patient's skin. With this arrangement, current passes from the active electrode through the patient's tissues to the  
10 external return plate. Since the patient represents a significant portion of the circuit, input power levels have to be high (typically 150 to 250 watts), to compensate for the resistive current limiting of the patient's tissues and, in the case of underwater electrosurgery, power losses due to the fluid medium which is rendered partially conductive by the presence of blood or other body fluids. Using high power with a  
15 monopolar arrangement is also hazardous, due to the tissue heating that occurs at the return plate, which can cause severe skin burns. There is also the risk of capacitive coupling between the instrument and patient tissues at the entry point into the body cavity.

20 With bipolar electrosurgery, a pair of electrodes (an active electrode and a return electrode) are used together at the tissue application site. This arrangement has advantages from the safety standpoint, due to the relative proximity of the two electrodes so that radio frequency currents are limited to the region between the electrodes. However, the depth of effect is directly related to the distance between the  
25 two electrodes; and, in applications requiring very small electrodes, the inter-electrode spacing becomes very small, thereby limiting tissue effect and the output power. Spacing the electrodes further apart would often obscure vision of the application site, and would require a modification in surgical technique to ensure direct contact of both electrodes with the tissue.

There are a number of variations to the basic design of the bipolar probe. For example, U.S. Patent Specification No. 4706667 describes one of the fundamentals of the design, namely that the ratio of the contact areas of the return electrode and of the active electrode is greater than 7:1 and smaller than 20:1 for cutting purposes. This range  
5 relates only to cutting electrode configurations. When a bipolar instrument is used for desiccation or coagulation, the ratio of the contact areas of the two electrodes may be reduced to approximately 1:1 to avoid differential electrical stresses occurring at the contact between the tissue and the electrode.

10 The electrical junction between the return electrode and tissue can be supported by wetting of the tissue by a conductive solution such as normal saline. This ensures that the surgical effect is limited to the active electrode, with the electric circuit between the two electrodes being completed by the tissue. One of the obvious limitations with the design is that the active electrode (typically a needle) must be completely buried in the  
15 tissue to enable the return electrode to complete the circuit. Another problem is one of the orientation: even a relatively small change in application angle from the ideal perpendicular contact with respect to the tissue surface, will change the contact area ratio, so that a surgical effect can occur in the tissue in contact with the return electrode.

20 Cavity distension provides space for gaining access to the operation site, to improve visualisation, and to allow for manipulation of instruments. In low volume body cavities, particularly where it is desirable to distend the cavity under higher pressure, liquid rather than gas is more commonly used due to better optical characteristics, and  
25 because it washes blood away from the operative site.

Conventional underwater electrosurgery has been performed using a non-conductive liquid (such as 1.5% glycine) as an irrigant, or as a distension medium to eliminate electrical conduction losses. Glycine is used in isotonic concentrations to prevent  
30 osmotic changes in the blood when intra-vascular absorption occurs. In the course of an operation, veins may be severed, with resultant infusion of the liquid into the

circulation, which could cause, among other things, a dilution of serum sodium which can lead to a condition known as water intoxication.

5 The applicants have found that it is possible to use a conductive liquid medium, such as normal saline, in underwater endoscopic electrosurgery in place of non-conductive, electrolyte-free solutions. Normal saline is the preferred distension medium in underwater endoscopic surgery when electrosurgery is not contemplated, or a non-electrical tissue effect such as laser treatment is being used. Although normal saline (0.9%w/v; 150mmol/l) has an electrical conductivity somewhat greater than that  
10 of most body tissue, it has the advantage that displacement by absorption or extravasation from the operative site produces little physiological effect, and the so-called water intoxication effects of non-conductive, electrolyte-free solutions are avoided.

15 Carbon dioxide is the preferred gaseous distension medium, primarily because of its non-toxic nature and high water solubility.

The applicants have developed a bipolar instrument suitable for underwater electrosurgery using a conductive liquid or gaseous medium. This electrosurgical  
20 instrument for the treatment of tissue in the presence of a fluid medium, comprises an instrument body having a handpiece and an instrument shaft and an electrode assembly, at one end of the shaft. The electrode assembly comprises a tissue treatment (active) electrode which is exposed at the extreme distal end of the instrument, and a return electrode which is electrically insulated from the tissue treatment electrode and has a  
25 fluid contact surface spaced proximally from the exposed part of the tissue treatment electrode. In use of the instrument, the tissue treatment electrode is applied to the tissue to be treated whilst the return electrode, being spaced proximally from the exposed part of the tissue treatment electrode, is normally spaced from the tissue and serves to complete an electrosurgical current loop from the tissue treatment electrode through the  
30 tissue and the fluid medium. This electrosurgical instrument is described in the specification of our International Patent Application No. PCT/GB96/01473.

The electrode structure of this instrument, in combination with an electrically-conductive fluid medium largely avoids the problems experienced with monopolar or bipolar electrosurgery. In particular, input power levels are much lower than those generally necessary with a monopolar arrangement (typically 100 watts). Moreover,  
 5 because of the relatively large spacing between its electrodes, an improved depth of effect is obtained compared with conventional bipolar arrangements.

The specification of our International Patent Application No. GB96/01472 describes an irrigated bipolar electrosurgical instrument that can be used in open air or gas-filled  
 10 environments. This instrument includes an internal channel for feeding electrically-conductive fluid (typically saline) to the exposed end of a tissue treatment electrode so as to provide a conductive fluid path that completes an electrical circuit to a return electrode when the instrument is in use. This instrument also includes an internal  
 15 channel for removing fluid from the region of the exposed end of the tissue treatment electrode. When the fluid is a liquid, such as saline, the presence of that liquid can cause collateral tissue damage, so its removal is desirable. This type of instrument is intended primarily for use in open air or gas-filled environments, and is not suitable for use with electrosurgical procedures which require distension of a body cavity.

20 However, where the volume of a body cavity is small - for example in arthroscopic surgery where even the large joints, such as the knee, may only accommodate 50-60 ml of irrigation fluid - the following problems may occur, namely:-

- 25 (i) Heated fluid in the immediate vicinity of the tissue contact electrode can cause collateral tissue damage;
- (ii) The products of the tissue vaporised by the tissue contact electrode can cause visualisation problems; and
- (iii) Soft tissue present in a joint space tends to move about, making it difficult to apply the active electrode to vaporise such tissue.

An arthroscope electrode may be characterised as short (100 to 140 mm), and rigid with a working diameter up to 5 mm. It can be introduced through a stab incision into a joint cavity (with or without a cannula) using the triangulation technique. Such an electrode is operated with a motion which moves the electrode between the 9 O'Clock and 3 O'Clock positions on the arthroscopic image. As a result, the tissue to be treated is usually approached at a shallow working angle with respect to the axis of the electrode. An arthroscopic electrode thus needs to have an effect consistent with this angled approach to the tissue. The tissue to be treated, such as meniscal cartilage, is commonly dense and of a high electrical impedance. An arthroscope electrode requires output power and voltage settings that reflect the type of tissue being treated, the size of electrode, and the fact that arthroscopists are seeking a speed of effect comparable to that of the mechanical shaver devices they currently employ, albeit with an electrode of smaller dimensions than a shaver blade for improved access.

The specification of our British Patent Application 9612993.7 describes an electrosurgical instrument for the treatment of tissue in the presence of an electrically-conductive fluid medium. The instrument comprises an instrument shaft, and an electrode assembly at one end of the shaft, the electrode assembly comprising a tissue treatment electrode and a return electrode which is electrically insulated from the tissue treatment electrode by means of an insulation member. The tissue treatment electrode has an exposed end for treating tissue, and the return electrode has a fluid contact surface which is spaced from the tissue treatment electrode in such a manner as to define, in use, a conductive fluid path that completes an electrical circuit between the tissue treatment electrode and the return electrode. The electrode assembly is provided with a plurality of apertures in the region of the tissue treatment electrode, through which apertures vapour bubbles and/or particulate material can be aspirated from the region surrounding the tissue treatment electrode.

An RF generator is provided for powering the electrode assembly. The power required from the RF generator to achieve vaporisation depends on a number of variables more fully described in the specification of our International Patent Application No.

GB97/00065. Of these variables two, are of particular importance in the context of the present invention; one being the cooling effect produced by the aspiration of conductive fluid in the region of the tissue contact electrode, and the other being the disruption of the vapour pocket formed around the tissue contact electrode by the flow of conductive fluid. These problems can be partially overcome by co-ordinating the aspiration by monitoring the output features of the generator which indicate the vaporisation power threshold has been exceeded. This usually results in a series of suction pulses as the vaporisation threshold is repeatedly exceeded between pulses and then elevated during the suction pulses so that, should vaporisation be maintained, the suction will be applied continuously. By using this technique, heated saline in the vicinity of the tissue contact electrode and vaporisation products can be successfully removed. The other desirable feature is the aspiration of loose tissue towards the tissue contact electrode, so that it can be stabilised during vaporisation. Whilst this can be achieved according to this technique; there are two significant performance limitations.

The first of these limitations is that the gaseous products of tissue vaporisation contain fatty products which have a sublimation property, i.e. they condense directly to a solid; sublimation occurring at temperatures well above boiling point. As the electrode shaft within the body cavity is cooled by the surrounding saline, these products are easily condensed. Thus, if a parallel suction shaft is used, the build up is along its entire length, and eventually completely blocks the tube. This process, even at the flow rates dictated by minimal influence on the power threshold, can cause very rapid blocking. For example, it is found that, with a moderately large electrode tip, using a 1mm internal diameter suction tube, complete blockage occurs after 30 seconds of activation. Obviously, a larger tube bore would increase the time before blockage, but this occurs so rapidly that the required bore size for a useful electrode life is beyond the dimensions of the maximum shaft diameter. The problems of sublimation are compounded by aspiration of tissue pieces which are incompletely vaporised before being excised from the remainder of the tissue. Given the need to attract tissue and, therefore, the requirement for a strong suction pressure which, once tissue is engaged with the tissue contact electrode and the vaporisation threshold is continually exceeded by cessation of



flow, increases the propensity for aspiration of unvaporised tissue and blockage of the aspiration channel.

The second of these limitations also relates to adherence of tissue to the tissue contact electrode. As indicated above, once the tissue obstructs flow, the vaporisation power threshold is exceeded, and suction is continuously applied. Under these circumstances, and particularly when aspiration channels are provided adjacent to the tissue treatment electrode, a steady state can be reached wherein the tissue is held around the periphery of the tissue contact electrode, the portion of tissue in the immediate vicinity of the tissue treatment electrode is vaporised but, without moving the application site or redirecting suction solely through the tissue treatment electrode, no further removal of tissue will occur. For example, large pieces of tissue tend to bridge the tissue treatment electrode, so that all tissue in contact with the electrode is removed, but the bulk of the tissue is left in place. Applying suction solely through the tissue treatment electrode limits the size of the electrode otherwise two extremes are created where, on the one hand during activation in conductive fluid, the vaporisation power threshold is very elevated despite synchronising suction pulses with the RF output, typically > 200 Watts, yet, on the other hand, can be reduced to below 50% of this level once tissue is engaged. With a static tissue contact electrode, there is an inevitable compromise between these performances variables.

The aim of the invention to provide an improved electrosurgical instrument of this type.

The present invention provides an electrosurgical instrument for the treatment of tissue in the presence of an electrically-conductive fluid medium, the instrument comprising an instrument shaft, and a tissue treatment electrode mounted at the distal end of the shaft for electrosurgically excising tissue pieces at an operation site, the tissue treatment electrode being movable cyclically relative to the distal end of the shaft, wherein vaporisation threshold control means are provided for controlling the power threshold for vaporisation of the electrically-conductive fluid medium at the tissue treatment electrode.

Advantageously, the instrument further comprises a return electrode which is electrically insulated from the tissue treatment electrode by insulation means, the tissue treatment electrode being exposed at the distal end of the instrument, and the return electrode having a fluid contact surface spaced proximally from the exposed end of the  
5 tissue treatment electrode.

Advantageously, the tissue treatment electrode is provided at the distal end of a rod mounted within, and movable relative to, the instrument shaft. Conveniently, the tissue treatment electrode is constituted by the distal end portion of the rod.

10

In a preferred embodiment, the instrument further comprises drive means for reciprocating the rod within the channel. Advantageously, the drive means is constituted by a motor and coupling means for converting the rotary output of the motor into reciprocatory movement of the rod.

15

In another preferred embodiment, the instrument further comprises drive means for rotating the rod within the channel. An electric motor may constitute the drive means.

In yet another preferred embodiment, the control means is such as to produce turbulence  
20 in the fluid medium surrounding the tissue treatment electrode, whereby, should vaporisation of the fluid medium surrounding the tissue treatment electrode cease, the control means is effective to reduce the turbulence, thereby to allow vaporisation to be re-established.

25 Preferably, the control means is such as to vary the rate of cyclical movement of the tissue treatment electrode thereby producing said turbulence in the fluid medium surrounding the tissue treatment electrode.

The instrument may further comprise removal means, the instrument having an  
30 apertured portion through which matter can be aspirated by the removal means from the

region surrounding the tissue treatment electrode, the removal means comprising a channel formed within the instrument shaft and leading from the apertured portion.

Preferably, the removal means further comprises a pump connected to the channel at  
5 a region thereof remote from the apertured portion of the instrument. In this case, the control means may be such as to vary the rate of the pump thereby producing said turbulence in the fluid medium surrounding the tissue treatment electrode. In this case, the pump may be activated cyclically whereby matter is aspirated by the removal means in a pulsed fashion. Conveniently, the pump is activated only when the tissue  
10 treatment electrode is powered for tissue vaporisation.

The instrument may further comprise an RF generator having a bipolar output connected to the tissue treatment electrode and the return electrode. Preferably, the control means includes a vaporisation detection circuit for detecting vaporisation in the region of the  
15 tissue treatment electrode, the vaporisation detection circuit being connected to the output of the RF generator whereby the vaporisation power threshold is controlled in dependence upon the output characteristics of the RF generator.

The control means may further include an RF active detection circuit for detecting the  
20 presence of RF voltages at the output of the RF generator, whereby the pump can be activated before the RF generator is turned on.

The electrosurgical instrument of the invention is useful for dissection, resection, vaporisation, desiccation and coagulation of tissue, as well as for combinations of these  
25 functions. It has a particular application in arthroscopic surgery as it pertains to endoscopic and percutaneous procedures performed on joints of the body including, but not limited to, such techniques as they apply to the spine and other non-synovial joints. Arthroscopic operative procedures may include: partial or complete meniscectomy of the knee joint including meniscal cystectomy; lateral retinacular release of the knee  
30 joint; removal of anterior and posterior cruciate ligaments or remnants thereof; labral tear resection, acromioplasty, bursectomy and subacromial decompression of the

shoulder joint; anterior release of the temporomandibular joint; synovectomy, cartilage debridement, chondroplasty, division of intra-articular adhesions, fracture and tendon debridement as applied to any of the synovial joints of the body; inducing thermal shrinkage of joint capsules as a treatment for recurrent dislocation, subluxation or  
 5 repetitive stress injury to any articulated joint of the body; discectomy either in the treatment of a disc prolapse or as part of a spinal fusion via a posterior or anterior approach to the cervical, thoracic and lumbar spine or any other fibrous joint for similar purposes; excision of diseased tissue; and haemostasis.

- 10 The instrument of the invention is also useful for dissection, resection, vaporisation, desiccation and coagulation of tissue, as well as combinations of these functions, with particular application in urological endoscopic (urethroscopy, cystoscopy, ureteroscopy and nephroscopy) and percutaneous surgery. Urological procedures may include: electro-vaporisation of the prostate gland (EVAP) and other variants of the procedure  
 15 commonly referred to as transurethral resection of the prostate (TURP) including, but not limited to, interstitial ablation of the prostate gland by a percutaneous or perurethral route whether performed for benign or malignant disease; transurethral or percutaneous resection of urinary tract tumours as they may arise as primary or secondary neoplasms, and further as they may arise anywhere in the urological tract from the calyces of the  
 20 kidney to the external urethral meatus; division of strictures as they may arise at the pelviureteric junction (PUJ), ureter, ureteral orifice, bladder neck or urethra; correction of ureterocoele; shrinkage of bladder diverticular; cystoplasty procedures as they pertain to corrections of voiding dysfunction; thermally induced shrinkage of the pelvic floor as a corrective treatment for bladder neck descent; excision of diseased tissue; and  
 25 haemostasis.

The electrosurgical instrument of the invention is also useful for dissection, resection, vaporisation, desiccation and coagulation of tissue and combinations of these functions with particular application in laparoscopic, colposcopic (including vaginal speculum)  
 30 and open surgical procedures on the female genital tract and adnexal related diseases. Laparoscopic operative procedures may include: removal of subserosal and pedunculated

fibroids, ablation of ectopic endometrium, ovarian cystectomy and ovarian drilling procedures; oophorectomy; salpingo-oophorectomy, subtotal hysterectomy and laparoscopically assisted vaginal hysterectomy (LAVH) as may be performed for benign or malignant diseases; laparoscopic uterosacral nerve ablation (LUNA); fallopian  
 5 tube surgery as correction of ectopic pregnancy or complications arising from acquired obstructions; division of abdominal adhesions; and haemostasis.

The electrosurgical instrument of the invention is also useful in the lower female genital tract, including treatment of cervix, vagina and external genitalia whether accessed  
 10 directly or using instrumentation comprising generally speculae and colposcopes. Such applications include: vaginal hysterectomy and other pelvic procedures utilising vaginal access; LLETZ/LEEP procedure (large loop excision of the transformation zone) or excision of the transformation zone of the endocervix; removal of cystic or septic lesions; ablation of genital or venereal warts; excision of benign and malignant lesions;  
 15 cosmetic and surgical repairs including vaginal prolapse; excision of diseased tissue; and haemostasis.

The electrosurgical instrument of the invention is also useful for dissection, resection, vaporisation, desiccation and coagulation of tissue and combinations of these functions  
 20 with particular application in surgery on the ear, nose and throat (ENT), and more particularly procedures performed on the oropharynx, nasopharynx and sinuses. These procedures may be performed through the mouth or nose using speculae or gags or using endoscopic techniques such as functional endoscopic sinus surgery (FESS). Functional endoscopic sinus procedures may include: removal of chronically-diseased  
 25 inflamed and hypertrophic mucus linings, polyps and neoplasms from the various anatomical sinuses of the skull; excision of diseased tissue; and haemostasis. Procedures on the nasopharynx may include: removal of chronically-diseased inflamed and hypertrophic mucus linings, polyps and neoplasms from the turbinates and nasal passages; submucous resection of the nasal septum; excision of diseased tissue; and  
 30 haemostasis. Procedures on the oropharynx may include: removal of chronically-diseased, inflamed and hypertrophic tissue, polyps and neoplasms particularly as they

occur related to the tonsil, adenoid, epi-glottic and supra-glottic regions, and salivary glands; as an alternative method to perform the procedure commonly known as laser assisted uvulopalatoplasty (LAUP); excision of diseased tissue; and haemostasis.

- 5 It is evident from the scope of applications of the invention that it has further additional applications for dissection, resection, vaporisation, desiccation and coagulation of tissue and combinations of these functions in general laparoscopic, thoracoscopic and neurosurgical procedures, being particularly useful in the removal of diseased tissue and neoplastic disease whether benign or malignant.

10

- Surgical procedures using the electrosurgical instrument of the invention may also include introducing the electrode assembly to the surgical site, whether through an artificial conduit (a cannula) or a natural conduit, which may be in an anatomical body cavity or space, or one created surgically. The cavity or space may be distended during the procedure using a fluid, or may be naturally held open by anatomical structures.
- 15 The surgical site may be bathed in a continuous flow of conductive fluid such as saline solution either to fill and distend the cavity, or to create a locally-irrigated environment around the tip of the electrode assembly in a gas filled cavity. The irrigating fluid may be aspirated from the surgical site to remove products created by application of the RF
- 20 energy, tissue debris or blood. The procedures may include simultaneous viewing of the site via an endoscope, or using an indirect visualisation means. An irrigated bipolar electrosurgical instrument is described in the specification of our International Patent Application No. PCT/GB96/01472.

- 25 The invention will now be described in greater detail, by way of example with reference to the drawings, in which:-

Figure 1 is a diagram showing an electrosurgical apparatus constructed in accordance with the invention;

- Figure 2 is a diagrammatic side elevation, partially broken away, of a first form of
- 30 electrode unit constructed in accordance with the invention;

Figure 3 is a diagrammatic side elevation of the electrode assembly of the electrode unit of Figure 2;

Figure 4 is a block circuit diagram illustrating vaporisation power threshold control of the electrode unit of Figures 2 and 3;

- 5 Figure 5 is a diagrammatic side elevation, partially broken away, of a second form of electrode unit constructed in accordance with the invention;

Figure 6 is a diagrammatic side elevation of the electrode assembly of the electrode unit of Figure 5;

- 10 Figure 7 is a diagrammatic side elevation, partially broken away, of a third form of electrode unit constructed in accordance with the invention;

Figure 8 is a diagrammatic side elevation of the electrode assembly of the electrode unit of Figure 7;

Figure 9 is a diagrammatic side elevation, partially broken away, of a fourth form of electrode unit constructed in accordance with the invention; and

- 15 Figure 10 is a diagrammatic side elevation of the electrode assembly of the electrode unit of Figure 9.

Referring to the drawings, Figure 1 shows electrosurgical apparatus including a generator 1 having an output socket 2 providing a radio frequency (RF) output, via a  
 20 connection cord 4, for an instrument in the form of a handpiece 3. Activation of the generator 1 may be performed from the handpiece 3 via a control connection (not shown) in the cord 4, or by means of a footswitch unit 5 connected separately to the rear of the generator 1 by a footswitch connection cord 6. In the illustrated embodiment, the footswitch unit 5 has two footswitches 5a and 5b for selecting a  
 25 desiccation mode and a vaporisation mode of the generator 1 respectively. The generator front panel has push buttons 7a and 7b for respectively setting desiccation and vaporisation power levels, which are indicated in a display 8. Push buttons 9 are provided as an alternative means for selection between the desiccation and vaporisation modes.

The handpiece 3 mounts a detachable electrode unit E, such as the electrode units E1 and E4 to be described below.

Figure 2 shows the first form of electrode unit E1 for detachable fastening to the  
 5 electrosurgical instrument handpiece 3, the electrode unit comprising a shaft 10, which  
 is constituted by a tube made of stainless steel. A tissue treatment (active) electrode  
 12 is provided at the distal end portion of the shaft 10. The active electrode 12 is  
 provided by the distal end portion of a rod 14 made of tungsten, the active electrode  
 extending at right angles to the rod. The rod 14 has a diameter of 0.4 to 0.6 mm. A  
 10 ceramic tube 18 is fixed to the rod 14 immediately adjacent to the active electrode 12.  
 A ceramic tip 20 is fixed within the out-turned distal end portion of the shaft 10.

As shown in Figure 2, the active electrode 12 protrudes through a longitudinal slot 20a  
 formed in the ceramic tip 20. That portion of the rod 14 not covered by the ceramic  
 15 tube 18 is provided with an insulating sleeve 22 made of polyimide,  
 polytetrafluoroethylene or by separate sleeves made by these two substances. A heat  
 sleeve 24 made of polytetrafluoroethylene or polyimide, covers the adjoining regions  
 of the ceramic tube 18 and the sleeve 22.

20 The major portion of the length of the shaft 10 is provided with an insulating heat  
 shrink sleeve 26 made of polyvinylidene fluoride. The sleeve 26 does not cover the  
 distal end portion of the shaft 10, that region of the shaft constituting a return electrode  
 28.

25 The rod 14 is mounted for reciprocal movement within the shaft 10, that end of the rod  
 remote from the active electrode 12 being fixed to a coupling member 30 slidably  
 mounted within one end 32a of a sleeve 32 made of stainless steel. The other end 32b  
 of the sleeve 32 is fixed to the adjacent end portion of the shaft 10. A top hat washer  
 34 is located within the sleeve end 32b, the washer constituting a backing member for  
 30 a silicone gland 36 and a delrin bush 38. A return spring 40 acts between the bush 38



An off-set shaft 30a is fixed to the end face of the coupling member 30, the free end of this shaft being engageable with an inclined end face 42a of a rotatable coupling member 42 fixed to the rotary output shaft of a motor 44. Hence, rotation of the output shaft of the motor 44 results in reciprocation of the coupling member 30 and the  
 5 rod 14.

The hollow interior of the shaft 10 is connected to a transverse tubular member 10a which is connected to a suction pump (not shown), and so constitutes a suction/exhaust port. As shown in Figure 2, the active electrode 12 is positioned at the end of an  
 10 aspiration channel constituted by the annular cavity defined by the interior of the shaft 10 and the rod 14, so that vapour bubbles and/or particulate material which, in use, are formed in the region of the active electrode, can be aspirated from the region for removal via the slot 20a, the aspiration channel and the port 10a.

15 The RF generator 1 (not shown in Figure 2) delivers an electrosurgical current to the electrodes 12 and 28 via connectors 46 and 48 provided respectively on the coupling member 30 and on the sleeve 32. The generator 1 includes means for varying the delivered output power to suit different electrosurgical requirements. Thus, in a first output power range of from about 140 volts to 200 volts, the active electrode 12 is used  
 20 for tissue desiccation; and, in a second output power range of from about 250 volts to 600 volts, the active electrode is used for tissue removal by cutting or vaporisation. For both ranges, the voltages are peak voltages. The generator 1 may be as described in the specification of our European Patent Application 96304558.8.

25 This electrosurgical instrument is particularly useful for rapid tissue debulking and the removal of loose tissue. One of the problems which could be encountered when tissue is rapidly debulked using an arthroscopic electrode configuration, particularly when working in small joint spaces, is the production of vapour bubbles generated as an end product of tissue vaporisation. Such bubbles obscure vision, and can coalesce at the site  
 30 of tissue application, so that the electrical circuit between the active and return electrodes becomes compromised by the absence of conductive fluid. Irregular active

product of tissue vaporisation. Such bubbles obscure vision, and can coalesce at the site of tissue application, so that the electrical circuit between the active and return electrodes becomes compromised by the absence of conductive fluid. Irregular active electrodes having filamentary, mesh or coiled spring forms goes some way to solving this problem as it reduces the vaporisation threshold as disclosed in the specification of our International patent application No. GB97/00065.

- The provision of the suction pump ensures the elimination of vapour bubbles from an operation site, which is particularly advantageous during aggressive tissue debulking.
- 10 The suction pump is activated only when the active electrode 12 is powered for tissue vaporisation. The pump is, therefore, pulsed so as to pull saline over the active electrode 12 (and to extract vapour bubbles and/or particulate material). This cools the active electrode 12, resulting in the collapse of the vapour pocket surrounding the active electrode. This, in turn, leads to the suction pump being turned off, thereby reducing the flow of saline over the active electrode 12. This electrode 12 then heats up again, leading to the re-formation of a vapour pocket, and the re-activation of the suction pump. This cycle then repeats until the generator 1 is turned off when the instrument is removed from the operation site.
- 15
- 20 The suction pump must be controlled so that the flow of bubbles from the active electrode 12 is balanced to the output characteristics of the RF generator 1 to prevent excessive cooling of the active electrode and a resultant increase in its vaporisation power threshold. The thermal mass of the thin, wire-form active electrode 12 is lower than that of a standard solid form active electrode, and this assists in rapidly re-establishing the vapour pocket around the active electrode should this collapse following excessive cooling.
- 25

The electrode unit E1 is intended primarily for use in arthroscopic surgery which requires rapid tissue debulking by vaporisation. The side-effect electrode (i.e. where the treatment axis is perpendicular to the shaft) configuration of the unit E1 is particularly advantageous for this purpose. In use, the electrosurgical instrument is

30

manipulated to introduce the electrode assembly constituted by the active electrode 12 and the return electrode 28 into a selected operation site (e.g. within the joint space of a knee), so that the active electrode contacts the tissue to be treated, and the tissue and the electrode assembly are immersed in saline.

5

The footswitch 5b (or the push button 7b) is then operated to activate the generator 1. The generator 1 then provides sufficient RF power to the electrode assembly to vaporise the saline surrounding the active electrode 12, and to maintain a vapour pocket surrounding this electrode. Using a brushing technique, with firm pressure against the  
10 tissue surface, rapid debulking of the tissue is achieved. Gently touching the tissue will reduce the effect, and can be used to sculpture and smooth the residual tissue surface.

With tissue engagement, the flow of irrigant away from the active electrode 12 will be reduced, the amount of reduction depending on the nature of the tissue surface, the application pressure and the suction pressure. Speed of debulking will, therefore,  
15 depend on these variables. Once the vaporisation occurs, the products will include vapour bubbles, carbon particles and tissue debris. All of these products are removed from the region of the active electrode 12, via the shaft 10 and the port 10a, by the suction pump.

20 All the constituents removed from the active tip are at high temperatures. This could lead to a potentially dangerous heating of the electrode shaft 10, which could cause tissue damage at the entry point. It may be, therefore, necessary to aspirate additional coolant saline from the body cavity along the inside surface of the shaft. To ensure that this saline is indeed at a safe temperature, it is taken from the rear of the return  
25 electrode 28 via a mesh filter (not shown).

In use, when the generator 1 is turned on, the motor 44 begins to rotate, causing the rod 14 to oscillate with an amplitude of 0.5mm. The oscillation of the rod 14 within the shaft 10 provides a mechanical agitation within the shaft that is sufficient to dislodge  
30 any sublimation products which condense within the shaft. In this way, blockage of the shaft 10 is prevented, so that the instrument can be used on a continuous basis.

The oscillation of the active electrode 12 also ensures that tissue pieces removed electrosurgically by vaporisation from an operation side are morcellated electrosurgically by the oscillating electrode, thereby preventing large tissue pieces bridging the aspiration channel. Morcellation is the division of a tissue piece into many  
 5 smaller pieces in order to facilitate its surgical removal.

The electrode unit E1 is also very effective in removing heated saline (distension fluid) from within a joint cavity. The risk of hot distension fluid occurs primarily during power application to reach the vaporisation threshold. Once the threshold has been  
 10 reached, the power requirement falls by 30-50%.

Whilst aspiration from the region of the active electrode 12 will remove heated saline from the body cavity, and remove any risk of overheating through prolonged activation under conditions where the vaporisation threshold is not reached, the cooling effect and  
 15 disruption of vapour pockets created around the active electrode will increase the vaporisation threshold. A vicious cycle can, therefore, be created, wherein the more suction applied at the active electrode 12, the more power required to reach the vaporisation threshold, and the greater the risk of heating. The other factor influencing the vaporisation threshold is the ratio of return: active contact area, and the insulation  
 20 separation between the active electrode 12 and the return electrode 28. The size of the active electrode 12 and the insulation separation, should, therefore, be reduced to the minimum necessary to achieve the function in order to offset the effects of aspiration in elevating the power threshold of vaporisation.

25 The specification of our International Patent Application GB97/00065 discloses techniques for controlling the vaporisation threshold by employing active electrode designs which assist in capturing vapour pockets and preventing cooling of the active electrode application site by screening from the flow of irrigant provided by channels in an endoscope. An alternative method of reducing the vaporisation power threshold  
 30 is to pulse the suction pressure, thereby allowing the threshold to be attained between pulses. Such pulses may be synchronised with the output features of the RF generator

1 to provide power bursts during active suction to sustain the vapour pocket, and clear any tissue occluding the apertures in the active electrode 12.

5 A known technique in arthroscopic surgery is to apply suction through a mechanical, tissue-nibbling device so that soft tissue present in the joint space, such as the infrapatellar fat pad, can be held in position within the nibbler jaws by suction whilst it is progressively "nibbled away".

10 Attracting tissue to the active electrode 12 of the electrode unit E1 has a similar effect as, for the reasons already given above, compliant tissue adhering to the active electrode will result in a reduction of the vaporisation power threshold. Adherent tissue will be rapidly vaporised, and small tissue particles produced during vaporisation will be aspirated from the application site.

15 Because of its speed of debulking and side-effect configuration, the electrode unit E1 also has advantages in urological surgery as an EVAP technique for use in conjunction with a resectoscope. A resectoscope electrode unit is introduced very differently, in that it is mounted on an endoscope prior to passage of the assembled instrument through a working sheath via the urethra. The proximal end of the electrode unit is connected  
20 to a trigger assembly and an electrical contact which is integral with the resectoscope. By this means, the electrode unit E1 can be moved back and forth through a defined range of motion by operating the trigger mechanism. As the electrode unit E1 is assembled prior to introduction, the size of the tip is not constrained by working channel dimensions, but rather by the diameter of the working sheath which can be up  
25 to 10 mm. Part of this diameter is occupied by the support wires to the electrode unit E1, which wires are commonly bent in a downward angle, with respect to the endoscopic image, to the working tip, so that they do not interfere with either visulation or its operation. Because of the reciprocatory movement of the rod 14, the active electrode 12 operates over a length lying within the range of from 3 mm to 4 mm and  
30 a width lying in the range of from 2 mm to 3 mm, and this size is necessary for

urological surgery given that, on average, 20-30 grammes of prostate tissue must be removed.

Because of the reservoir effect of the urinary bladder, and the mounting of the  
5 endoscope to view the tip of the active electrode 12 from below, bubble generation  
during vaporisation is less of a problem during endoscopic urology, as the bubbles flow  
away from the endoscope to accumulate in the bladder. Nevertheless, the use of the  
electrode unit E1 substantially reduces the possibility of bubble generation causing  
problems.

10

Although the electrode unit E1 is intended primarily for use in the vaporisation of tissue  
it can also be used for desiccation, particularly of synovial membranes or to separate  
muscle attachments. In this case, once the electrode assembly of the electrode unit E1  
has been introduced into a selected operation site, the RF generator 1 is actuated using  
15 the footswitch 5a or the push button 7a. The generator 1 will then provide sufficient  
RF power to the electrode assembly to maintain the saline adjacent to the active  
electrode 12 substantially at its boiling point without creating a vapour pocket  
surrounding that electrode. The instrument can then be manipulated by moving the  
active electrode 12 across the surface of the tissue to be treated in a side-to-side  
20 "painting" technique.

The electrode unit E1 can also be used for delivering a blended power output. This is  
achieved by automatically alternating the output of the RF generator 1 between the  
desiccation and vaporisation power levels, more haemostasis being produced than is  
25 possible in the vaporisation mode. As a consequence, the speed of tissue debulking is  
reduced, but the increased haemostasis is useful when cutting or debulking vascular  
tissue structures. Alternatively, the output of the RF generator 1 can be pulsed at the  
vaporisation power level, without cycled activation of the desiccation mode. This  
produces a less aggressive tissue vaporisation than occurs in the vaporisation mode, with  
30 a consequent reduction in both bubble formation and the risk of tissue charring.

The active electrode 12 of the unit E1 is a side effect electrode (i.e. its treatment axis is perpendicular to the shaft). Axial agitation is advantageous with such electrodes, in that the entire electrode can be brought into contact with tissue. As a result, the exposed area can be made very small, allowing operation at lower powers and less at  
 5 higher saline flow rates.

Reference above has been made to pulsing the operation of the suction pump to control the vaporisation power threshold at the active electrode 12. It is also possible to influence the vaporisation power threshold by reciprocating the electrode 12 so as to  
 10 produce turbulence in the surrounding saline solution, the reciprocation speed being controlled or being pulsed depending on the degree to which the vaporisation power threshold is to be altered.

Referring to Figure 4, an electronic controller for achieving either power vaporisation  
 15 threshold control function has a drive unit 100 for driving a pump motor 104, and a drive unit 102 for driving the motor 44 in response to variable pulse width modulation inputs derived from pulse width modulation circuits 108, 110. Both modulation circuits are controllable by a microcontroller 112 and a vaporisation detection circuit 114. The microcontroller 112 is, itself, under the control of an RF output active  
 20 detection circuit 116 and a footswitch detection circuit 118. The maximum speeds of the two motors 104, 44 are user selected using switches 120, and are used to set up pulse width modulation pulse trains in each case by means of the respective circuit 108, 110. However, the pulse widths from each circuit 108, 110 are also controlled by the vaporisation detection circuit 114. The circuit has an input 114I connected to the  
 25 output of the RF generator 1, and contains peak voltage sensing components as disclosed in our European Patent Application No. 0754437. When vaporisation at the electrode 12 occurs, the electrode peak voltage rises rapidly. When vaporisation at the electrode 12 occurs, the electrode peak voltages rises rapidly. Detection of this rise by the vaporisation detection circuit 114 causes the mark-to-space ratio of the signals  
 30 fed to the motor drive circuits 100, 102 to increase, thereby increasing the vaporisation

power threshold so as to tend to reduce vaporisation. This arrangement can be set up so that the motors 104 and 44 remain inactive until vaporisation is detected.

In an alternative mode, the pump motor 104 can be operated before vaporisation is detected, and indeed, before the generator 1 is switched on. Thus, by detecting the presence of RF voltages at the output of the generator 1 using the RF active detection circuit 116, the pump can be activated by the microcontroller 112 for a preprogrammed duration in order to attract and remove loose tissue in the vicinity of the electrode 12 when the generator is switched on. However, by using the footswitch detection circuit 118 connected to the footswitch unit 5, and by controlling the actuation of the generator 1 by the microcontroller 112 via a line 126, a different operation sequence can be effected whereby, upon actuation of the footswitch unit, firstly the pump motor 104 is started before RF power is applied to the electrode 12, then a signal is fed to the generator via the line 126, after a predetermined interval, to actuate the output of the generator. This means that, instead of the pump controller being a slave unit to the RF generator 1 as first described above, the generator is a slave unit to the pump controller. Consequently, a boosted pump phase can be activated without RF power, so as to attract loose tissue before it is dispersed by the application of RF power.

Figures 5 and 6 show the second form of electrode unit E2. This instrument is a modification of that shown in Figures 2 and 3, and so like reference numerals will be used for like parts, and only the modifications will be described in detail. There are two main modifications, the first being to the drive to the rod 14, and the second to the configuration of the active electrode 12.

25

In the first modification, the motor 44 rotatably drives the rod 14 via a coupling assembly 42. As with the embodiment of Figures 2 and 3, the rod 14 passes through aligned apertures in the washer 34, the gland 36 and the delrin bush 38. The bush 38 is somewhat longer than the equivalent bush of the embodiment of Figures 2 and 3 extending to the end 32a of the sleeve 32. A slip ring 46a is provided to connect the connector 46 to the rod 14.



The other main modification is that the active electrode 12 (the free end of the tungsten rod 14 - in this embodiment of 0.5mm diameter) is bent back over the free end of the ceramic tube 18. The turned-back portion 12a of the electrode 12 constitutes a side effect electrode. An apertured region 20a is formed between the ceramic tip 20 and the active electrode 12, this region leading to the aspiration channel defined by the interior of the shaft 10.

Another modification is that the rod 14 is a flexible drive rod whose distal end portion is off-set with respect to the central longitudinal axis of the shaft 10. In use, when the generator 1 is turned on, the motor 44 begins to rotate, causing the rod 14 to rotate within the shaft 10. This rotation provides a mechanical agitation that is sufficient to dislodge any sublimation products which condense within the shaft. The off-set of the rod 14 results in an unstable oscillation being set up in the rod, which sweeps adherent tissue debris from the inner wall of the shaft 10.

Figures 7 and 8 show the third form of electrode unit E3. This unit E3 is a modification of the unit E2, so like reference numerals will be used for like parts, and only the modifications will be described in detail. The main modification is to the configuration of the active electrode assembly. Thus, as shown in Figure 8, the active electrode 12 is shaped like a crank handle, and defines an elbow 12b which is off-set from the axis of the ceramic tube 18. The ceramic tip 20 is formed with an inclined cam surface 20b which, in use, engages with the elbow 12b to force the tip of the active electrode 12 outwardly, and to ensure better tissue engagement. This crank handle configuration of the active electrode 12 also ensures that, as the tip rotates, the elbow 12b is pushed around the inner surface of the ceramic tip 20, thereby removing debris which would otherwise tend to build up there.

Figures 9 and 10 show the fourth form of electrode unit E4. This unit E4 is also a modification of the unit E2, so like reference numerals will be used for like parts, and only the modifications will be described in detail. Here, the main modification is to the configuration of the active electrode 12 which, in this case, is an end effect electrode,

being constituted by a simple hook-shaped end portion 12a at the end of the rod 14. As with the embodiments of Figures 5 and 6, the rod 14 is a flexible drive rod whose distal end portion off-set with respect to the central longitudinal axis of the shaft 10.

- 5 As with the electrode unit E1 of Figures 2 and 3 the suction pump of each of the electrode units E2 to E4 can be pulsed to control the vaporisation power threshold at the active electrode 12. The control circuit of Figure 4 could be used for this purpose, the circuit in each of these cases being used to influence the vaporisation power threshold by rotating (instead of reciprocating) the active electrode 12 so as to produce turbulence  
 10 in the surrounding saline solution, the rotation speed being controlled or being pulsed depending upon the degree to which the vaporisation threshold is to be altered.

As has already been described, the adherence of tissue over the active electrode 12 may induce a steady state condition, and the aspiration method must allow for removal of  
 15 unvaporised tissue particles whilst not quenching vapour pocket formation. Rotation of the active electrode 12 of the electrode units E2 to E4 provides several advantages to overcome these performance issues. Thus, rotating the active electrode 12 increases the effective size of the electrode, as far as tissue contact area is concerned, for one complete rotation, whilst reducing the physical size of the active electrode. Reducing  
 20 the size of the active electrode 12 reduces the vaporisation power threshold to a degree sufficient to enable aspiration along the axis of rotation when the generator control method is employed.

The introduction of rotation and aspiration through the active electrode 12, or more  
 25 accurately through a channel within the range of motion of the active electrode, prevents the steady state being reached, and so prevents tissue bridging. This is achieved as tissue temporarily obstructing the aspiration channel is always treated, as opposed to positioning aspiration channels outside the range of motion of the active electrode 12, in which case only tissue adjacent to that obstructing the aspiration channel would be  
 30 treated.

Given that the aspiration channel is required to cope with unvaporised tissue, the active electrode 12 is only required to incise the tissue such that the tip of the tissue in the aspiration channel is detached from the body of the tissue and then aspirated through the channel. Ideally, the truncated portion of tissue is also morcellated or partially vaporised by the active electrode 12 to reduce the size of tissue pieces. This morcellation is accomplished by introducing an off-set in the drive shaft/connector to the active electrode 12 which rotates in the aspiration channel of larger internal diameter than the external diameter of the connector, a feature which has additional advantages in preventing blocking of the aspiration channel, as is described below.

10

The relative contributions of tissue incision or morcellation and tissue vaporisation to the overall tissue debulking process can be controlled by the interaction of the bore of the terminal aspiration channel, the suction pressure and the bulk of the active electrode 12. Owing to the overall size constraints on the external diameter of the instrument it is, in general, the diameter of the drive rod 14 whose distal tip forms the active electrode 12 and which, therefore, also provides the means of electrical connection to the active electrode, which determines whether tissue removal occurs primarily by incision/morcellation or vaporisation. Typically a drive rod 14 (and hence active electrode 12) formed from 0.2-1.0 mm diameter tungsten wire provides incision/morcellation, and a drive rod active electrode formed from 0.5mm diameter tungsten wire primarily provides vaporisation. The incision/morcellation technique has advantages when dealing with soft friable tissue, whereas the vaporisation technique has advantages when application is made to dense fibrous or cartilaginous tissue. The design can, therefore, be optimised for the type of tissue encountered during use in particular surgical specialities or, alternatively, a multi functional design with a drive rod and active electrode typically formed from 0.4-0.6mm tungsten can be used.

For all four electrode units E1 to E4, agitation within the aspiration shaft 10 significantly reduces the risk of blockage, either by morcellated tissue, sublimated products of vaporisation or both. This can be accomplished by axial or rotary motion of the rod 14 which is positioned within the aspiration channel, with or without other

30

means of fluid agitation, including the cycling of suction pressure, which may be provided as an integral feature of generator output, control of suction, and sonic pressure waves. To enhance the effect of agitation, it is beneficial to construct the drive rod 14 from a lubricious material to reduce adherence.

5

Each of the electrode units E1 and E4, has the additional advantage that the aspiration in the region of the active electrode 12 restricts the flow of convection currents in the saline surrounding the electrode assembly. As the power threshold required to reach vaporisation is dependent on the power dissipation of the active electrode 12 and the  
 10 flow characteristics around it, the power threshold is dependent upon the maximum rate of convection. Consequently, the restriction of the convection currents reduces the power threshold and/or permits the use of higher saline flow rates, and this is advantageous as it enables the use of a cheaper RF generator, as well as avoiding problems such as dissipation within the instrument, and catastrophic overheating of the  
 15 active electrode. It also facilitates control of the generator once vaporisation commences. The importance of power threshold of vaporisation is discussed in greater detail in the specification of our International Patent Application No. GB97/00065.

20 Moreover, each of the electrode units, E1 to E4 is such as to prevent tissue bridging, as the tendency for tissue to obstruct the aspiration channel is, in each case, obviated by the movement of the active electrode ensuring that such tissue is treated. The movement of the active electrode 12 also ensures tissue morcellation, though this is effected by electrosurgery rather than by mechanical cutting.

25

It will be apparent that modifications could be made to the electrode units described above. For example, instead of providing an off-set drive rod 14, this rod could be loosely coiled so that the coils lie against the inner wall of the aspiration channel, whereby, during rotation, a worm screw action occurs to encourage proximal  
 30 movement of tissue debris, as well as cleaning of the inner wall of the channel.

It would also be possible to introduce axial motion during rotation. Thus, for the electrode unit E4, the simple 90° hook form active electrode 12 can rotate on a bearing surface provided by the distal end face of the ceramic tube 18, this end face being provided with ratchet teeth features. Thus, as the rod 14 rotates, the hook-shaped end portion 12a moves in and out as it engages and disengages the ratchet teeth, this axial movement being permitted by the off-set flexible drive rod 14 repeatedly elongating and shortening.

As an alternative to an electric motor, each of the units E1 to E4 could be powered by a fluid drive generated through a rotary vane or similar apparatus, which, in turn, may be powered by the suction means.

It is also possible to power the rotary drive by the RF generator 1, so that an integral and interactive system of the rotary drive, the active electrode 12, the RF generator and the suction means is provided.

The upper limit of the speed of rotation of the units E2 to E4 is defined at that level which elevates the vaporisation power threshold beyond the output range of the RF generator 1, which will, in turn, be dependent upon the geometry of the active electrode 12. Typically, the speed of tissue removal is increased with increased rotary speed when primarily employing the incision/morcellation technique, and is increased with decreased rotary speed when primarily employing the vaporisation technique. It is, therefore, evident that, in a multi-functional design, it is advantageous for the user to vary the rotary speed depending on the nature of the tissue being treated. To this end, a typical range of rotary speeds would be from 100 revs/min to 1000 revs/min.

With the rotary action electrode units E2 to E4, the effective size of the active electrode 12 is increased, and a significant aspect is the incision of tissue. The active electrode 12 is fabricated from the distal end of the drive rod 14, so simple wire form electrodes meet these performance requirements. The only drawback of these simple electrode forms is that asymmetry of the tissue contact can make it difficult to maintain an

accurate location on a tissue surface, particularly when that surface is comprised of more fibrous or more dense tissue.

If the wire form active electrode 12 protrudes from the ceramic tube 18, for example  
 5 in a simple loop form as with the electrode unit E2, then the potential exists for the loop  
 to excise tissue pieces too large for aspiration through the distal opening of the  
 aspiration channel. Should this occur, the exposed distal end of the drive rod 14  
 within the aspiration channel performs an important function in morcellating and  
 vaporising such tissue pieces, so that they are reduced in size sufficiently to enter the  
 10 aspiration channel. This function is enhanced by the eccentric motion of the drive rod  
 14 within the aspiration channel.

Whilst the amount of protrusion of the active electrode 12 from the distal end of the  
 ceramic tube 18 is governed by the rules described in our International Patent  
 15 Application GB96/01473, the effect of aspiration in increasing vaporisation threshold  
 changes these rules. The other performance factor governing the dimension of the  
 active electrode 12 is similar to that defining the diameter of the wire. Thus, the  
 thinner wire forms, which are used on soft tissue, can protrude from the distal end of  
 the ceramic tube 18 in the treatment axis; whilst the thicker wire forms, which are  
 20 used on more dense tissue, ideally extend beyond the distal end of the ceramic tube in  
 the treatment axis by an amount not exceeding the diameter of the wire.

The active electrode 12 may also take on more convoluted or more complex generally  
 planar forms of end effect electrodes and generally axial forms for side effect  
 25 electrodes, for example coils, spirals, meshes or multiple spokes.

Our International Patent Application GB96/01472 describes a technique of introducing  
 a conductive fluid to the region of a tissue treatment (active) electrode in order to  
 define, in use, a conductive fluid path between the active electrode and a return  
 30 electrode. The electrode units E1 to E4 of the present invention could be modified to

incorporate those features. In particular, these units could be modified for use in gaseous operating environments, either on the surface of a body or within body cavities.

The specification of our British Patent Application 9612993.7 describes a technique of aspiration in the vicinity of a tissue treatment (active) electrode, wherein the suction pressure is controlled by generator output features in order to facilitate vaporisation by intermittently lowering the vaporisation threshold by cessation of suction flow. The techniques could advantageously be incorporated in the electrode units E1 to E4, both to ensure the vaporisation threshold is exceeded between suction pulses, and as a result of the suction pulsing assisting in preventing blockage of the aspiration channel.

As a suction pulse is initiated only once the vaporisation threshold has been exceeded, tissue can only be attracted to the active electrode once the threshold is exceeded by activation remote from the tissue within the surrounding distension medium. It is known that the vaporisation threshold is lowered once tissue is engaged by the active electrode. It is, therefore, advantageous for suction to be applied initially without RF activation as a variable time delay feature.

In summary the electrosurgical instrument of the invention has the following advantageous features:-

1. A small active electrode surface which is able to treat large tissue areas by virtue of active electrode movement.
2. A small active electrode to enable vaporisation, despite the cooling effects created by aspiration.
3. A mechanical movement at the active electrode tip, compatible with material removal within the aspiration channel.
4. Aspiration operation is dependent upon the vaporisation condition.
5. At least the outside of the shaft 10 is coated with a non-stick material such as polytetrafluoroethylene - ideally the inside of the shaft as well.

6. Active electrode tip movement occurs across the face of the aspiration channel, so that any lodged tissue is electrosurgically morcellated.
7. Active electrode agitation is dependent upon the vaporisation condition.
8. Discontinuities within the agitator rod ensure that the internal surfaces of the shaft are cleaned: or the rod flexes sufficiently to create the same effect.
- 5 9. A ceramic-to-ceramic interface at the active electrode tip ensures that the internal circumference of the outer ceramic is wiped by the inner ceramic.
10. The agitator rod is independently insulated in ceramic at its tip.
11. Offset rotary action for a side-effect electrode to enable flat surface engagement.



CLAIMS

1. An electrosurgical instrument for the treatment of tissue in the presence of an  
5 electrically-conductive fluid medium, the instrument comprising an instrument shaft,  
and a tissue treatment electrode mounted at the distal end of the shaft for  
electrosurgically excising tissue pieces at an operation site, the tissue treatment electrode  
being movable cyclically relative to the distal end of the shaft, wherein vaporisation  
threshold control means are provided for controlling the power threshold for  
10 vaporisation of the electrically-conductive fluid medium at the tissue treatment  
electrode.
2. An electrosurgical instrument as claimed in claim 1, further comprising a return  
electrode which is electrically insulated from the tissue treatment electrode by  
15 insulation means, the tissue treatment electrode being exposed at the distal end of the  
instrument, and the return electrode having a fluid contact surface spaced proximally  
from the end of the tissue treatment electrode.
3. An electrosurgical instrument as claimed in claim 1 or claim 2, wherein the  
20 tissue treatment electrode is provided at the distal end of a rod mounted within, and  
movable relative to, the instrument shaft.
4. An electrosurgical instrument as claimed in claim 3, wherein the tissue treatment  
electrode is constituted by the distal end portion of the rod.
- 25 5. An electrosurgical instrument as claimed in claim 3 or claim 4, further  
comprising drive means for reciprocating the rod within the channel.
6. An electrosurgical instrument as claimed in claim 5, wherein the drive means  
30 is constituted by a motor and coupling means for converting the rotary output of the  
motor into reciprocatory movement of the rod.

7. An electrosurgical instrument as claimed in claim 3 or claim 4, further comprising drive means for rotating the rod within the channel.

8. An electrosurgical instrument as claimed in claim 7, wherein an electric motor  
5 constitutes the drive means.

9. An electrosurgical instrument as claimed in any one of claims 1 to 8, wherein the control means is such as to produce turbulence in the fluid medium surrounding the tissue treatment electrode, whereby, should vaporisation of the fluid medium  
10 surrounding the tissue treatment electrode cease, the control means is effective to reduce the turbulence, thereby to allow vaporisation to be re-established.


10. An electrosurgical instrument as claimed in claim 9, wherein the control means is such as to vary the rate of cyclical movement of the tissue treatment electrode thereby  
15 producing said turbulence in the fluid medium surrounding the tissue treatment electrode.



11. An electrosurgical instrument as claimed in any one of claims 1 to 10, further comprising removal means, the instrument having an apertured portion through which  
20 matter can be aspirated by the removal means from the region surrounding the tissue treatment electrode, the removal means comprising a channel formed within the instrument shaft and leading from the apertured portion.

12. An electrosurgical instrument as claimed in claim 11, wherein the removal  
25 means further comprises a pump connected to the channel at a region thereof remote from the apertured portion of the instrument.

13. An electrosurgical instrument as claimed in claim 12 when appendant to claim  
30 9, wherein the control means is such as to vary the rate of the pump thereby producing said turbulence in the fluid medium surrounding the tissue treatment electrode.

- 14     An electrosurgical instrument as claimed in claim 12 or claim 13, wherein the pump is activated cyclically whereby matter is aspirated by the removal means in a pulsed fashion.
- 5   15.     An electrosurgical instrument as claimed in any one of claims 12 to 14, wherein the pump is activated only when the tissue treatment electrode is powered for tissue vaporisation.
- 10   16.     An electrosurgical instrument as claimed in claim 2, or any one of claims 3 to 15 when appendant to claim 2, further comprising an RF generator having a bipolar output connected to the tissue treatment electrode and the return electrode.
- 15   17.     An electrosurgical instrument as claimed in claim 16 when appendant to any one of claims 10 to 15, wherein the control means includes a vaporisation detection circuit for detecting vaporisation in the region of the tissue treatment electrode, the vaporisation detection circuit being connected to the output of the RF generator whereby the vaporisation power threshold is controlled in dependence upon the output characteristics of the RF generator.
- 20   18.     An electrosurgical instrument as claimed in claim 17 when appendant to any one of claims 12 to 16, wherein the control means further includes an RF active detection circuit from detecting the presence of RF voltages at the output of the RF generator, whereby the pump can be activated before the RF generator is turned on.










Fig.4.






Fig. 7.






Fig. 8.







Application No: GB 9715199.7  
Claims searched: 1 to 18

Examiner: David Brunt  
Date of search: 30 October 1997

**Patents Act 1977**  
**Search Report under Section 17**

**Databases searched:**

UK Patent Office collections, including GB, EP, WO & US patent specifications, in:

UK Cl (Ed.O): A5R (RHCC, RHCE)

Int Cl (Ed.6): A61B (17/36, 17/38, 17/39)

Other: Online: EDOC, JAPIO, WPI

**Documents considered to be relevant:**

Category	Identity of document and relevant passage	Relevant to claims
A	GB 2308981 A (GYRUS) note page 19 line 25	-
A	EP 0754437 A1 (GYRUS) note Fig 4	-

X Document indicating lack of novelty or inventive step  
Y Document indicating lack of inventive step if combined with one or more other documents of same category.  
& Member of the same patent family

A Document indicating technological background and/or state of the art.  
P Document published on or after the declared priority date but before the filing date of this invention.  
E Patent document published on or after, but with priority date earlier than, the filing date of this application.

**This Page is Inserted by IFW Indexing and Scanning  
Operations and is not part of the Official Record**

**BEST AVAILABLE IMAGES**

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

- ☐ BLACK BORDERS
- ☐ IMAGE CUT OFF AT TOP, BOTTOM OR SIDES
- ☐ FADED TEXT OR DRAWING
- ☐ BLURRED OR ILLEGIBLE TEXT OR DRAWING
- ☐ SKEWED/SLANTED IMAGES
- ☐ COLOR OR BLACK AND WHITE PHOTOGRAPHS
- ☐ GRAY SCALE DOCUMENTS
- ☒ LINES OR MARKS ON ORIGINAL DOCUMENT
- ☐ REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY
- ☐ OTHER: \_\_\_\_\_

**IMAGES ARE BEST AVAILABLE COPY.**

**As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.**

**This Page Blank (uspto)**